

# **SPECIFICATION**

# 64-02-015-PDS

Page 1 of 2 2/12/2014

## NFPA 99 Compliant Duplex Medical Air Desiccant Dryer System 15 SCFM with Regulators, Safety Relief Valves and Dew Point/CO Monitor

#### **Duplex Dryer System**

- Two desiccant twin tower air dryers with an integral demand based purge saving control system
- Two pre-filters
- Two after filters
- Two final line regulators and safety relief valves
- Four valve bypass for dryer maintenance
- Isolation valves to permit servicing without shutting down the medical air system
- Air sample port

#### Dryer

- Twin tower, pressure swing adsorption, regenerative type
- Each dryer individually sized for peak calculated demand and capable of producing a 10° F (-12° C) pressure dew point
- Desiccant is contained within a packed bed canister to prevent excessive downstream dusting
- Purge flow is minimized through an integral purge saving control system

#### **Pre-filter**

- High efficiency coalescing pre-filter rated for 0.01 micron mounted on each dryer with automatic drain and element change indicator
- Automatic solenoid drain valve for coalescing filter

#### After Filter

 Particulate final line filters rated for 0.01 micron mounted on each dryer with element change indicator

## **Regulator/ Safety Relief Valve**

- Dual regulators
- Factory set @ 55 PSI with
- Safety relief valves rated for 75 PSI.

## Dew Point Hygrometer/CO Monitor

- Dew point hygrometer/CO monitor with integral chemical CO sensor - mounted, pre-piped, wired and includes remote alarm contacts
- Hygrometer sensor ceramic type
- Accuracy minimum of ± 2° F for dew point and ± 2 PPM (at 10 PPM) for carbon monoxide
- Dew point alarm factory set at 36° F (2° C) per NFPA 99
- CO alarm factory set at 10 PPM
- Set points are not field adjustable
- High CO and high dew point conditions indicated with visual and audible alarms.

#### **Control System**

- Mounted and wired control system
- Allows the operator to choose between manual purge and demand purge
- In manual purge, dryer will shift towers every 2 minutes
- Online tower re-pressurizes for 50 seconds
- 15% of system capacity used for purge air
- In demand purge, the dew point monitor controls the dryer purge as purging depends on dew point condition

# **SPECIFICATION**

## 64-02-015-PDS



Page 2 of 2 2/12/2014



Maximum Working Pressure: 125 psig

## Maximum Inlet Temperature: 110° F

## Statement of Warranty

Patton's Medical warrants all Medical Air Dryer packages, to be free of defects in material and workmanship under normal use for a period not to exceed thirty (30) months from date of shipment, or twenty-four (24) months from date of start-up.